

Department	Equipment	Equipment Vendor	JSN #	PP New Reuse	Installation Location	Testing & Certification Requirements Amount of time per item, per discipline				Concurrent Testing		Dates	NOTES
						MTF BMET	KTR BMET	Physicist	Other	Yes	No		
<b>ENT</b>													
Final Move	OR Table	Berchtold	M9080	Reuse	1C572		1 hr	N/A				01/13/21	
Final Move	Mini-Cat	TBD	U0003	Reuse	1C573	4 hrs	1 hr	5 days				01/11-01/15/21	
<b>Musculo Clinic</b>													
Chiropractic	Laser Stim Unit	Multi Radiance Medical	U0084	Reuse			.5 hr	N/A				12/09/20	
Orthopedics	Surgical Lights		M7470	New	1C238.1		.5 hr	N/A				12/10/20	
Orthopedics	GE XR656	GE Healthcare	X5900	New	1C238.1	2 days	1 hr	3 days		X		12/09-12/11/20	
<b>Radiology</b>													
North Angio	Exam Lights (Ceiling Mounted)		M7435	Reuse	Prep Recovery		.5 hr	N/A				02/05/21	
North Angio	Surgical Lights		M7470	Reuse	2B600		.5 hr	N/A				02/05/21	
North Angio	Gentlemax Laser		U9014	New		.5 hr	.5 hr	1 day		X		02/05/21	
North Angio	Cardiac UltraSound LOGIQE9	GE Healthcare	X2100	New		4 hrs	1 hr	N/A				02/05/21	
North Angio	Toshiba Biplane Angio Cath Lab	Toshiba	X6175	New	2B600	2 days	1 hr	3 days			X	02/01-02/05/21	
Radiography	GE Discovery XR650	GE Healthcare	X5900	New	2B620	2 days	1 hr	3 days			X	02/01-02/05/21	
Bone Density	Hologic Discovery A	Hologic	X2120	New	2B659.1	2 days	1 hr	3 days				TBD	
S. Angio	Toshiba Single Plane Angio Cath Lab		X6175	New	2B611	2 days	1 hr	5 days			X	TBD	
Mammo	Hologic Selenia Dimensions	Hologic	X5430	New	2B665.3	2 days	1 hr	3 days				TBD	

**Instruction:** Identify all new and reuse property items that require testing and certification beyond standard safety and operational checks, personnel performing certification (i.e. Contractor BMET, Govt BMET, Physicist, Third-party Vendor, etc.) and task duration per unit. Include existing property relocating to both interim and final spaces. Provide in Excel format and include information found on this example document.